
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001847
10:16 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:42
STERILIZING 3:51
136.4 C 224kPa 3:51
136.4 C 225kPa 4:21
136.4 C 225kPa 4:51
136.4 C 225kPa 5:21
136.3 C 226kPa 5:51
136.4 C 226kPa 6:21
136.4 C 226kPa 6:51
136.4 C 225kPa 7:21
136.4 C 225kPa 7:21
VENTING 7:21
AIR DRYING 8:20
CYCLE COMPLETE 14:59

Digital Signature #
E92B9E7250C209E2
