
Midmark M11 - v1.0.0
Total Cycles: 2343

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2022 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 197.8 F 4.5
4:00 206.8 F 10.0
6:00 224.3 F 5.6
8:00 237.9 F 11.6
10:00 236.3 F 9.3
12:00 237.8 F 8.9
14:00 250.3 F 15.5
16:00 263.1 F 22.8
18:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.8 F 29.2
2:00 272.0 F 28.5
3:00 271.7 F 28.7
4:00 272.6 F 29.1

Min 271.3 F 28.2
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.2 F 0.0
5:00 213.2 F 0.0
10:00 205.4 F 0.0
15:00 211.9 F 0.0
20:00 211.0 F 0.0
25:00 226.4 F 0.0
30:00 235.9 F 0.0
Drying Complete

FILLING: 2:46
HEATING: 18:04
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

09/20/2022 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____