
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006887
10:58 01/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:48
STERILIZING 3:52
136.5 C 220kPa 3:52
136.4 C 225kPa 4:22
136.2 C 219kPa 4:52
136.2 C 221kPa 5:22
136.4 C 221kPa 5:52
136.2 C 223kPa 6:22
135.9 C 220kPa 6:52
136.3 C 222kPa 7:22
136.3 C 222kPa 7:22
VENTING 7:22
AIR DRYING 8:20
CYCLE COMPLETE 9:04

Digital Signature #
1E6B8B0D95A3205F
