
Midmark M11 - v1.0.5
Total Cycles: 1671

Name: RITTER2
Sterilizer ID: V2313228
11/02/2022 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.0 F 0.0
2:00 161.3 F 2.9
4:00 172.8 F 6.7
6:00 199.9 F 11.4
8:00 215.5 F 5.7
10:00 229.0 F 10.9
12:00 230.2 F 7.5
14:00 243.6 F 13.5
16:00 243.1 F 12.3
18:00 255.7 F 18.7
20:00 266.7 F 25.6
22:00 273.0 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.4
1:00 272.1 F 29.6
2:00 272.2 F 29.8
3:00 272.2 F 29.8
4:00 272.2 F 29.7

Min 272.0 F 29.3
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 214.3 F 0.0
10:00 238.1 F 0.0
15:00 235.3 F 0.0
20:00 221.9 F 0.0
25:00 215.3 F 0.0
30:00 210.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:42
CYCLE COMPLETE

11/02/2022 03:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____