
Midmark M11 - v1.0.0
Total Cycles: 798

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2021 08:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.7 F 0.0
2:00 193.6 F 2.9
4:00 196.5 F 7.7
6:00 215.1 F 11.4
8:00 225.8 F 8.4
10:00 236.7 F 10.5
12:00 241.2 F 12.0
14:00 239.7 F 10.5
16:00 254.0 F 17.7
18:00 266.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.1
2:00 271.5 F 28.6
3:00 272.6 F 29.2
4:00 271.5 F 28.4

Min 271.3 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 231.2 F 0.0
10:00 226.9 F 0.0
15:00 227.8 F 0.0
20:00 226.5 F 0.0
25:00 228.1 F 0.0
30:00 233.8 F 0.0

Drying Complete

FILLING: 1:39
HEATING: 19:27
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

04/22/2021 09:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____