
Midmark M11 - v1.0.5
Total Cycles: 312

Name: RITTER2
Sterilizer ID: V2313228
08/25/2021 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.8 F 0.0
2:00 142.8 F 2.2
4:00 179.0 F 6.5
6:00 195.2 F 11.6
8:00 215.3 F 5.5
10:00 227.1 F 10.5
12:00 231.7 F 7.6
14:00 244.3 F 13.3
16:00 243.3 F 12.3
18:00 255.9 F 19.0
20:00 267.2 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 272.7 F 29.2
1:00 272.5 F 29.7
2:00 272.2 F 29.7
3:00 272.1 F 29.2
4:00 272.5 F 29.9

Min 271.8 F 29.0
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.1 F 0.0
5:00 215.3 F 0.0
10:00 225.3 F 0.0
15:00 226.4 F 0.0
20:00 227.9 F 0.0
25:00 231.4 F 0.0
30:00 215.6 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:17
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

08/25/2021 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____