

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
[david@herzig-eye.com](mailto:david@herzig-eye.com)  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005399  
9:44 21/04/2023

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:29  
STERILIZING 3:30  
136.5 C 220kPa 3:30  
136.3 C 222kPa 4:00  
136.1 C 223kPa 4:30  
136.4 C 225kPa 5:00  
136.3 C 220kPa 5:30  
136.1 C 222kPa 6:00  
136.4 C 225kPa 6:30  
136.4 C 224kPa 7:00  
136.4 C 224kPa 7:00  
VENTING 7:01  
AIR DRYING 7:49  
CYCLE COMPLETE 10:21

-----

Digital Signature #  
70EFEBBE8EF46A70

-----