

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3107

Name: RITTER2  
Sterilizer ID: V2313228  
11/09/2023 02:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.0 F 0.0  
2:00 136.4 F 1.2  
4:00 160.3 F 4.1  
6:00 182.3 F 8.0  
8:00 199.3 F 12.4  
10:00 217.8 F 6.5  
12:00 227.7 F 11.0  
14:00 229.1 F 7.6  
16:00 241.8 F 13.3  
18:00 242.2 F 12.1  
20:00 254.4 F 18.5  
22:00 265.5 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.3  
1:00 272.6 F 30.0  
2:00 271.2 F 29.2  
3:00 272.3 F 30.2  
4:00 272.0 F 29.7

Min 271.0 F 29.2  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.2 F 0.0  
5:00 219.7 F 0.0  
10:00 211.3 F 0.0  
15:00 201.7 F 0.0  
20:00 196.3 F 0.0  
25:00 192.6 F 0.0  
30:00 193.5 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:46  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:02:26  
CYCLE COMPLETE

11/09/2023 03:46 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_