
Midmark M11 - v1.0.0
Total Cycles: 5226

Name: RITTER 1
Sterilizer ID: V2178198
01/08/2025 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.1 F 0.0
2:00 183.8 F 2.3
4:00 189.3 F 6.7
6:00 210.6 F 11.9
8:00 221.3 F 6.2
10:00 234.3 F 11.3
12:00 233.4 F 8.2
14:00 242.0 F 11.1
16:00 245.9 F 13.1
18:00 258.4 F 19.8
20:00 269.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 271.5 F 28.1
2:00 271.6 F 28.1
3:00 271.7 F 28.1
4:00 272.0 F 28.3

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 201.8 F 0.0
10:00 202.4 F 0.0
15:00 201.6 F 0.0
20:00 205.0 F 0.0
25:00 213.5 F 0.0
30:00 209.6 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

01/08/2025 11:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____