
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001269
13:10 22/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:25
PRESSURIZING 4:25
STERILIZING 5:36
136.0 C 221kPa 5:36
136.4 C 224kPa 6:06
136.4 C 225kPa 6:36
136.4 C 225kPa 7:06
136.4 C 226kPa 7:36
136.4 C 225kPa 8:06
136.4 C 224kPa 8:36
136.4 C 225kPa 9:06
136.4 C 225kPa 9:06
VENTING 9:06
AIR DRYING 10:03
CYCLE COMPLETE 14:43

Digital Signature #
82A0A7E24FF1B59D
