
Midmark M11 - v1.0.5
Total Cycles: 1008

Name: RITTER2
Sterilizer ID: V2313228
04/20/2022 06:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.3 F 0.0
2:00 75.7 F 0.0
4:00 89.9 F 0.0
6:00 110.8 F 0.0
8:00 163.0 F 0.1
10:00 187.3 F 0.3
12:00 206.8 F 1.4
14:00 216.6 F 4.7
16:00 226.5 F 8.1
18:00 235.9 F 11.9
20:00 233.5 F 8.5
22:00 244.3 F 12.6
24:00 243.4 F 12.3
26:00 241.9 F 11.3
28:00 254.2 F 17.6
30:00 265.1 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 29.1
2:00 271.7 F 28.8
3:00 272.4 F 29.0
4:00 271.6 F 28.7

Min 271.6 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 192.9 F 0.0
10:00 189.3 F 0.0
15:00 188.0 F 0.0
20:00 201.6 F 0.0
25:00 210.8 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 31:53
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

04/20/2022 07:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____