
Midmark M11 - v1.0.0
Total Cycles: 3146

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2023 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.0 F 0.0
2:00 192.2 F 0.0
4:00 209.8 F 4.7
6:00 224.6 F 9.8
8:00 230.9 F 8.4
10:00 241.6 F 12.2
12:00 239.9 F 10.5
14:00 238.1 F 9.5
16:00 253.5 F 16.9
18:00 265.7 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.1 F 28.4
2:00 272.1 F 28.7
3:00 271.7 F 28.1
4:00 272.4 F 28.8

Min 271.3 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 220.2 F 0.0
10:00 237.0 F 0.0
15:00 228.4 F 0.0
20:00 222.7 F 0.0
25:00 221.7 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:38
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

04/21/2023 02:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____