
Midmark M11 - v1.0.5
Total Cycles: 3576

Name: RITTER2
Sterilizer ID: V2313228
05/01/2024 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.5 F 0.0
2:00 151.6 F 2.4
4:00 162.7 F 5.8
6:00 187.0 F 10.1
8:00 211.2 F 5.7
10:00 222.7 F 8.2
12:00 234.8 F 13.0
14:00 235.9 F 10.2
16:00 235.9 F 8.7
18:00 248.9 F 15.3
20:00 260.8 F 22.0
22:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.9 F 29.9
2:00 271.9 F 29.2
3:00 271.4 F 29.4
4:00 272.9 F 29.9

Min 271.2 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 211.2 F 0.0
10:00 197.3 F 0.0
15:00 195.9 F 0.0
20:00 194.2 F 0.0
25:00 192.4 F 0.0
30:00 189.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

05/01/2024 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____