
Midmark M11 - v1.0.5
Total Cycles: 1346

Name: RITTER2
Sterilizer ID: V2313228
07/26/2022 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.4 F 0.0
2:00 177.2 F 0.4
4:00 188.7 F 4.4
6:00 208.4 F 9.6
8:00 223.2 F 7.3
10:00 234.1 F 11.2
12:00 234.5 F 9.2
14:00 238.9 F 9.6
16:00 248.7 F 15.3
18:00 261.2 F 22.3
20:00 272.4 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.5 F 29.7
2:00 271.6 F 29.2
3:00 272.6 F 29.9
4:00 272.0 F 29.2

Min 271.6 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 197.9 F 0.0
10:00 202.1 F 0.0
15:00 201.1 F 0.0
20:00 201.8 F 0.0
25:00 206.9 F 0.0
30:00 209.4 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

07/26/2022 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____