
Midmark M11 - v1.0.0
Total Cycles: 4466

Name: RITTER 1
Sterilizer ID: V2178198
04/17/2024 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.0 F 0.0
2:00 159.8 F 2.3
4:00 174.4 F 5.7
6:00 193.7 F 9.7
8:00 216.5 F 8.5
10:00 225.2 F 6.6
12:00 234.8 F 10.8
14:00 232.0 F 7.2
16:00 244.1 F 12.3
18:00 240.7 F 10.4
20:00 252.4 F 16.2
22:00 262.8 F 22.2
24:00 272.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.5
2:00 272.5 F 28.9
3:00 271.3 F 28.3
4:00 272.4 F 28.6

Min 271.2 F 28.2
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 195.6 F 0.0
10:00 207.2 F 0.0
15:00 206.8 F 0.0
20:00 215.7 F 0.0
25:00 235.8 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:10
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:04:01
CYCLE COMPLETE

04/17/2024 11:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____