
Midmark M11 - v1.0.5
Total Cycles: 4336

Name: RITTER2
Sterilizer ID: V2313228
02/05/2025 05:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.0 F 0.0
2:00 115.6 F 1.4
4:00 135.4 F 4.3
6:00 167.1 F 8.7
8:00 194.7 F 8.6
10:00 215.3 F 5.9
12:00 229.0 F 10.6
14:00 230.4 F 7.2
16:00 243.7 F 13.3
18:00 243.0 F 12.0
20:00 256.3 F 19.0
22:00 268.2 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.6 F 29.1
2:00 272.9 F 29.6
3:00 271.9 F 28.8
4:00 272.8 F 29.2

Min 271.8 F 28.6
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.4 F 0.0
5:00 209.9 F 0.0
10:00 215.2 F 0.0
15:00 215.4 F 0.0
20:00 213.1 F 0.0
25:00 208.9 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

02/05/2025 06:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____