
Midmark M11 - v1.0.0
Total Cycles: 1940

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2022 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 84.1 F 0.0
4:00 88.9 F 0.0
6:00 124.9 F 0.5
8:00 156.3 F 2.4
10:00 177.8 F 5.2
12:00 196.4 F 8.7
14:00 211.1 F 12.5
16:00 223.7 F 5.6
18:00 235.8 F 9.8
20:00 239.2 F 9.7
22:00 241.8 F 11.4
24:00 237.8 F 9.3
26:00 252.0 F 16.2
28:00 263.7 F 23.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.2 F 28.4
2:00 272.4 F 28.6
3:00 272.7 F 28.8
4:00 272.8 F 29.0

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 233.6 F 0.0
Dry Cycle Aborted

FILLING: 0:56
HEATING: 29:52
STERILIZING: 4:00
VENTING: 2:40
DRYING: 4:56
TOTAL CYCLE: 00:42:24

05/16/2022 12:37 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____