

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5315

Name: RITTER 1  
Sterilizer ID: V2178198  
02/05/2025 09:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.4 F 0.0  
2:00 69.7 F 0.4  
4:00 91.3 F 2.1  
6:00 128.3 F 4.6  
8:00 159.2 F 7.5  
10:00 173.7 F 10.6  
12:00 186.5 F 6.5  
14:00 206.9 F 4.3  
16:00 219.1 F 7.7  
18:00 229.9 F 11.5  
20:00 228.5 F 6.7  
22:00 241.2 F 12.0  
24:00 238.3 F 9.6  
26:00 251.6 F 16.1  
28:00 263.0 F 22.6  
30:00 272.2 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 272.0 F 28.5  
2:00 272.2 F 28.7  
3:00 272.4 F 28.9  
4:00 272.5 F 29.0

Min 271.7 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 202.1 F 0.0  
10:00 210.1 F 0.0  
15:00 208.9 F 0.0  
20:00 212.0 F 0.0  
25:00 212.0 F 0.0  
30:00 213.2 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 30:06  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:08:15  
CYCLE COMPLETE

02/05/2025 10:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_