
Midmark M11 - v1.0.0
Total Cycles: 3926

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2023 01:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 180.3 F 3.6
4:00 186.6 F 7.9
6:00 211.5 F 12.8
8:00 225.0 F 7.1
10:00 234.0 F 12.3
12:00 235.1 F 9.2
14:00 239.1 F 9.7
16:00 246.9 F 13.8
18:00 258.9 F 20.3
20:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.7 F 29.3
2:00 271.9 F 28.7
3:00 271.2 F 28.6
4:00 272.7 F 29.3

Min 271.2 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 220.9 F 0.0
10:00 244.6 F 0.0
15:00 236.7 F 0.0
20:00 225.4 F 0.0
25:00 225.2 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

10/19/2023 02:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____