

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003820  
11:40 15/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:31  
STERILIZING 3:21  
136.0 C 221kPa 3:21  
136.4 C 225kPa 3:51  
136.4 C 220kPa 4:21  
136.2 C 218kPa 4:51  
136.2 C 220kPa 5:21  
136.4 C 221kPa 5:51  
136.4 C 225kPa 6:21  
136.4 C 225kPa 6:51  
136.4 C 225kPa 6:51  
VENTING 6:51  
CYCLE ABORTED 7:21  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 7:21  
-----

Digital Signature #  
683FB5002CA9AA1F  
-----