
Midmark M11 - v1.0.5
Total Cycles: 1167

Name: RITTER2
Sterilizer ID: V2313228
06/07/2022 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.4 F 0.0
2:00 185.4 F 4.1
4:00 201.2 F 9.8
6:00 215.9 F 6.3
8:00 231.7 F 11.1
10:00 234.4 F 9.0
12:00 238.0 F 9.8
14:00 248.7 F 15.1
16:00 261.2 F 22.2
18:00 272.3 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.7 F 29.9
2:00 272.1 F 29.9
3:00 272.8 F 29.9
4:00 272.0 F 29.8

Min 271.8 F 29.4
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 208.9 F 0.0
10:00 222.6 F 0.0
15:00 225.2 F 0.0
20:00 226.5 F 0.0
25:00 228.4 F 0.0
30:00 212.9 F 0.0
Drying Complete

FILLING: 2:38
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

06/07/2022 12:43 PM
ACCEPTED: X
REJECTED:
Operator: U02