

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001996  
7:37 10/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:35  
PRESSURIZING 3:38  
STERILIZING 4:46  
136.4 C 223kPa 4:46  
136.2 C 222kPa 5:16  
136.4 C 225kPa 5:46  
136.3 C 224kPa 6:16  
136.3 C 224kPa 6:46  
136.4 C 225kPa 7:16  
136.4 C 223kPa 7:46  
136.4 C 225kPa 8:16  
136.4 C 225kPa 8:16  
VENTING 8:16  
AIR DRYING 9:14  
CYCLE COMPLETE 10:05  
-----

Digital Signature #  
BAB18E4A8C344733  
-----