
Midmark M11 - v1.0.0
Total Cycles: 4516

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2024 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.5 F 0.0
2:00 166.8 F 3.2
4:00 180.6 F 6.9
6:00 202.8 F 11.2
8:00 216.7 F 4.1
10:00 229.1 F 8.2
12:00 238.9 F 12.7
14:00 238.3 F 9.8
16:00 238.6 F 9.5
18:00 248.6 F 14.3
20:00 259.9 F 20.6
22:00 270.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.7
2:00 271.3 F 28.1
3:00 272.4 F 28.8
4:00 271.8 F 28.2

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 205.5 F 0.0
10:00 205.9 F 0.0
15:00 193.7 F 0.0
20:00 196.3 F 0.0
25:00 198.4 F 0.0
30:00 185.2 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 22:53
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:01:59
CYCLE COMPLETE

05/02/2024 11:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____