
Midmark M11 - v1.0.5
Total Cycles: 313

Name: RITTER2
Sterilizer ID: V2313228
08/25/2021 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.2 F 0.0
2:00 160.1 F 2.9
4:00 170.5 F 6.7
6:00 194.9 F 11.4
8:00 216.8 F 6.0
10:00 228.6 F 11.1
12:00 233.8 F 8.2
14:00 242.2 F 11.8
16:00 245.2 F 13.2
18:00 256.9 F 19.6
20:00 267.5 F 26.3
22:00 273.2 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 272.0 F 30.3
2:00 273.2 F 30.5
3:00 273.1 F 30.2
4:00 272.0 F 29.9

Min 271.0 F 29.5
Max 273.5 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 234.4 F 0.0
10:00 228.0 F 0.0
15:00 226.9 F 0.0
20:00 187.2 F 0.0
25:00 180.1 F 0.0
30:00 177.2 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:14
CYCLE COMPLETE

08/25/2021 02:27 PM
ACCEPTED: X
REJECTED:
Operator: U02