
Midmark M11 - v1.0.5
Total Cycles: 2201

Name: RITTER2
Sterilizer ID: V2313228
03/30/2023 02:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.4 F 0.0
2:00 141.4 F 2.0
4:00 158.0 F 5.3
6:00 182.8 F 9.8
8:00 213.0 F 5.6
10:00 224.8 F 9.2
12:00 233.2 F 9.5
14:00 239.8 F 12.0
16:00 239.4 F 10.7
18:00 252.5 F 17.4
20:00 264.3 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.5
1:00 272.7 F 30.0
2:00 271.1 F 29.5
3:00 272.4 F 30.1
4:00 271.4 F 29.5

Min 271.0 F 29.4
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 220.6 F 0.0
10:00 225.2 F 0.0
15:00 229.7 F 0.0
20:00 219.4 F 0.0
25:00 211.7 F 0.0
30:00 208.8 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:55
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

03/30/2023 03:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____