
Midmark M11 - v1.0.5
Total Cycles: 3232

Name: RITTER2
Sterilizer ID: V2313228
12/19/2023 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.6 F 0.0
2:00 108.7 F 0.0
4:00 105.6 F 0.6
6:00 132.9 F 2.9
8:00 165.9 F 6.0
10:00 174.9 F 9.8
12:00 195.2 F 13.6
14:00 213.8 F 6.1
16:00 226.0 F 10.0
18:00 231.4 F 9.7
20:00 237.0 F 11.0
22:00 234.4 F 8.6
24:00 247.2 F 14.5
26:00 257.7 F 20.3
28:00 267.4 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.0 F 29.8
2:00 271.9 F 29.4
3:00 271.6 F 29.6
4:00 272.3 F 29.7

Min 271.2 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 209.1 F 0.0
10:00 224.1 F 0.0
15:00 228.9 F 0.0
20:00 228.1 F 0.0
25:00 228.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:22
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:07:40
CYCLE COMPLETE

12/19/2023 12:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____