

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3541

Name: RITTER 1  
Sterilizer ID: V2178198  
07/18/2023 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
2:00 187.8 F 0.1  
4:00 185.7 F 2.7  
6:00 204.1 F 6.8  
8:00 219.3 F 11.8  
10:00 227.3 F 6.7  
12:00 238.0 F 11.7  
14:00 236.5 F 8.9  
16:00 239.7 F 10.0  
18:00 247.2 F 13.7  
20:00 259.3 F 20.2  
22:00 269.9 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.4 F 28.8  
2:00 271.7 F 28.1  
3:00 271.7 F 28.5  
4:00 272.4 F 28.7

Min 271.1 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
5:00 235.2 F 0.0  
10:00 217.3 F 0.0  
15:00 223.3 F 0.0  
20:00 216.5 F 0.0  
25:00 223.9 F 0.0  
30:00 234.4 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 22:58  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:10  
CYCLE COMPLETE

07/18/2023 10:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_