
Midmark M11 - v1.0.0
Total Cycles: 5220

Name: RITTER 1
Sterilizer ID: V2178198
01/06/2025 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.8 F 0.0
2:00 70.9 F 0.0
4:00 86.9 F 0.0
6:00 122.7 F 0.0
8:00 148.5 F 0.1
10:00 172.4 F 0.2
12:00 191.4 F 0.4
14:00 203.8 F 2.2
16:00 213.6 F 5.1
18:00 221.8 F 7.7
20:00 230.2 F 10.8
22:00 231.5 F 8.2
24:00 238.1 F 10.4
26:00 237.6 F 9.1
28:00 246.1 F 13.2
30:00 243.8 F 11.8
32:00 254.4 F 17.3
34:00 263.9 F 22.9
36:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 271.9 F 28.2
2:00 272.3 F 28.6
3:00 272.7 F 28.8
4:00 272.7 F 28.7

Min 271.5 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.2 F 0.0
5:00 206.7 F 0.0
10:00 215.6 F 0.0
15:00 209.7 F 0.0
20:00 202.8 F 0.0
25:00 205.0 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 36:34
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:15:11
CYCLE COMPLETE

01/06/2025 11:13 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____