

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002548  
11:57 28/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:59  
STERILIZING 3:05  
136.4 C 224kPa 3:05  
136.3 C 225kPa 3:35  
136.4 C 226kPa 4:05  
136.3 C 225kPa 4:35  
136.4 C 225kPa 5:05  
136.1 C 223kPa 5:35  
136.4 C 227kPa 6:05  
136.3 C 225kPa 6:35  
136.3 C 225kPa 6:35  
VENTING 6:35  
AIR DRYING 7:30  
CYCLE COMPLETE 36:11  
-----

Digital Signature #  
9C67FC26CE9C0FE0  
-----