
Midmark M11 - v1.0.0
Total Cycles: 1801

Name: RITTER 1
Sterilizer ID: V2178198
04/06/2022 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.0 F 0.0
2:00 176.3 F 3.2
4:00 187.5 F 7.6
6:00 211.9 F 13.5
8:00 227.1 F 7.9
10:00 242.8 F 13.2
12:00 239.6 F 10.6
14:00 236.3 F 8.8
16:00 251.0 F 15.8
18:00 263.1 F 22.7
20:00 272.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.4 F 29.0
2:00 271.2 F 28.2
3:00 272.3 F 28.9
4:00 271.2 F 28.2

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 231.2 F 0.0
10:00 199.4 F 0.0
15:00 192.3 F 0.0
20:00 191.0 F 0.0
25:00 205.5 F 0.0
30:00 233.4 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:01
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

04/07/2022 07:12 AM
ACCEPTED: X
REJECTED:
Operator: U02