
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002158
7:37 22/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:30
PRESSURIZING 3:33
STERILIZING 4:46
136.0 C 221kPa 4:46
136.2 C 222kPa 5:16
136.4 C 225kPa 5:46
136.4 C 224kPa 6:16
136.4 C 224kPa 6:46
136.3 C 222kPa 7:16
136.4 C 223kPa 7:46
136.4 C 224kPa 8:16
136.4 C 224kPa 8:16
VENTING 8:16
AIR DRYING 9:11
CYCLE COMPLETE 39:11

Digital Signature #
F986C364294DA723
