
Midmark M11 - v1.0.5
Total Cycles: 97

Name: RITTER2
Sterilizer ID: V2313228
06/14/2021 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.4 F 0.0
2:00 174.4 F 2.9
4:00 180.0 F 7.0
6:00 200.4 F 12.1
8:00 216.4 F 6.7
10:00 229.5 F 11.8
12:00 233.9 F 9.5
14:00 238.0 F 9.8
16:00 248.3 F 15.1
18:00 260.2 F 21.7
20:00 270.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.2
1:00 272.5 F 30.0
2:00 271.6 F 29.1
3:00 272.4 F 29.9
4:00 271.8 F 29.2

Min 271.4 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 221.3 F 0.0
10:00 222.0 F 0.0
15:00 218.7 F 0.0
20:00 214.2 F 0.0
25:00 211.7 F 0.0
Dry Cycle Aborted

FILLING: 1:48
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:16
DRYING: 29:07
TOTAL CYCLE: 00:58:51

06/14/2021 03:26 PM
ACCEPTED: X
REJECTED:
Operator: U01