
Midmark M11 - v1.0.0
Total Cycles: 2995

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2023 10:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 176.5 F 1.9
4:00 192.0 F 5.2
6:00 210.7 F 9.7
8:00 223.3 F 7.2
10:00 228.8 F 9.1
12:00 237.2 F 10.9
14:00 239.3 F 11.4
16:00 237.8 F 9.6
18:00 251.5 F 16.3
20:00 263.4 F 23.2
22:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.5 F 29.1
2:00 271.6 F 28.8
3:00 272.5 F 29.1
4:00 271.4 F 28.6

Min 271.4 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.7 F 0.0
5:00 217.9 F 0.0
10:00 216.7 F 0.0
15:00 236.4 F 0.0
20:00 235.0 F 0.0
25:00 228.4 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

03/13/2023 11:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____