

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2922

Name: RITTER 1  
Sterilizer ID: V2178198  
02/21/2023 10:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
2:00 200.3 F 4.2  
4:00 207.2 F 9.6  
6:00 226.1 F 7.0  
8:00 234.9 F 10.9  
10:00 235.2 F 8.4  
12:00 240.2 F 10.1  
14:00 248.0 F 14.4  
16:00 261.1 F 21.6  
18:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.1  
2:00 271.7 F 28.4  
3:00 271.9 F 29.0  
4:00 272.5 F 29.1  
  
Min 271.4 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 223.5 F 0.0  
10:00 228.5 F 0.0  
15:00 230.8 F 0.0  
20:00 229.0 F 0.0  
25:00 227.5 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 18:26  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:48  
CYCLE COMPLETE

02/21/2023 11:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_