
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001790
7:32 13/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:58
STERILIZING 4:05
136.0 C 220kPa 4:05
136.1 C 222kPa 4:35
136.4 C 224kPa 5:05
136.4 C 224kPa 5:35
136.4 C 221kPa 6:05
135.9 C 220kPa 6:35
136.3 C 223kPa 7:05
136.4 C 224kPa 7:35
136.4 C 224kPa 7:35
VENTING 7:36
AIR DRYING 8:23
CYCLE COMPLETE 25:20

Digital Signature #
4F5DC79D24B32307
