
Midmark M11 - v1.0.5
Total Cycles: 3972

Name: RITTER2
Sterilizer ID: V2313228
09/25/2024 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 72.0 F 0.0
4:00 80.1 F 0.0
6:00 96.8 F 0.1
8:00 150.0 F 0.3
10:00 172.0 F 0.6
12:00 193.2 F 2.1
14:00 206.8 F 5.2
16:00 220.5 F 8.4
18:00 230.2 F 12.0
20:00 229.6 F 7.7
22:00 240.4 F 12.6
24:00 239.1 F 10.5
26:00 236.4 F 8.9
28:00 250.1 F 15.5
30:00 261.0 F 21.7
32:00 271.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.2
2:00 271.6 F 28.7
3:00 272.6 F 29.3
4:00 271.6 F 28.6

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 190.5 F 0.0
10:00 202.9 F 0.0
15:00 205.7 F 0.0
20:00 208.2 F 0.0
25:00 215.4 F 0.0
30:00 217.8 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:45
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:11:03
CYCLE COMPLETE

09/25/2024 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____