

-----  
Midmark M11 - v1.0.5  
Total Cycles: 806

Name: RITTER2  
Sterilizer ID: V2313228  
02/16/2022 08:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.7 F 0.0  
2:00 88.8 F 0.0  
4:00 91.4 F 0.0  
6:00 112.5 F 0.2  
8:00 161.3 F 2.5  
10:00 178.5 F 6.1  
12:00 195.4 F 10.0  
14:00 214.0 F 7.9  
16:00 222.1 F 7.6  
18:00 231.7 F 12.1  
20:00 234.0 F 9.0  
22:00 240.4 F 10.7  
24:00 246.2 F 13.6  
26:00 257.9 F 20.0  
28:00 268.5 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 271.7 F 28.9  
2:00 271.7 F 29.0  
3:00 271.7 F 29.0  
4:00 271.6 F 28.8

Min 271.6 F 28.7  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
5:00 235.3 F 0.0  
10:00 217.0 F 0.0  
15:00 217.5 F 0.0  
20:00 213.6 F 0.0  
25:00 215.5 F 0.0  
30:00 212.4 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 29:03  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:07:05  
CYCLE COMPLETE

02/16/2022 09:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_