
Midmark M11 - v1.0.5
Total Cycles: 3739

Name: RITTER2
Sterilizer ID: V2313228
06/26/2024 04:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.0 F 0.0
2:00 155.9 F 2.5
4:00 180.5 F 6.2
6:00 195.7 F 10.5
8:00 216.9 F 6.0
10:00 226.1 F 9.9
12:00 232.4 F 9.1
14:00 239.9 F 12.6
16:00 240.6 F 11.5
18:00 253.3 F 17.9
20:00 264.4 F 24.6
22:00 271.1 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 272.7 F 30.2
2:00 271.8 F 29.6
3:00 272.1 F 30.4
4:00 272.0 F 29.8

Min 270.9 F 29.2
Max 273.2 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 205.7 F 0.0
10:00 206.6 F 0.0
15:00 206.0 F 0.0
20:00 212.2 F 0.0
25:00 213.6 F 0.0
30:00 211.3 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

06/26/2024 05:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____