
Midmark M11 - v1.0.5
Total Cycles: 2276

Name: RITTER2
Sterilizer ID: V2313228
04/20/2023 11:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.5 F 0.0
2:00 187.5 F 3.7
4:00 196.2 F 8.7
6:00 218.2 F 9.0
8:00 229.7 F 10.1
10:00 234.0 F 7.9
12:00 241.8 F 11.6
14:00 245.7 F 13.6
16:00 259.0 F 20.7
18:00 270.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.1 F 29.3
2:00 272.2 F 29.7
3:00 271.6 F 29.0
4:00 272.3 F 29.7

Min 271.3 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 224.0 F 0.0
10:00 242.5 F 0.0
15:00 241.3 F 0.0
20:00 227.8 F 0.0
25:00 226.4 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 18:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

04/20/2023 12:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____