
Midmark M11 - v1.0.5
Total Cycles: 4476

Name: RITTER2
Sterilizer ID: V2313228
04/02/2025 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.1 F 0.0
2:00 74.6 F 0.1
4:00 101.2 F 0.8
6:00 127.5 F 2.7
8:00 150.6 F 5.3
10:00 168.1 F 8.1
12:00 201.7 F 11.0
14:00 221.7 F 4.7
16:00 216.3 F 5.8
18:00 220.5 F 9.1
20:00 228.7 F 12.9
22:00 230.6 F 8.8
24:00 241.6 F 13.5
26:00 241.2 F 11.5
28:00 251.5 F 17.0
30:00 261.2 F 22.6
32:00 270.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.9
1:00 271.7 F 30.2
2:00 271.7 F 30.2
3:00 271.7 F 30.1
4:00 272.0 F 31.1

Min 271.0 F 29.7
Max 272.5 F 31.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 193.6 F 0.0
10:00 189.7 F 0.0
15:00 197.3 F 0.0
20:00 192.2 F 0.0
25:00 192.9 F 0.0
30:00 193.0 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:46
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:11:23
CYCLE COMPLETE

04/02/2025 10:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____