
Midmark M11 - v1.0.5
Total Cycles: 3076

Name: RITTER2
Sterilizer ID: V2313228
11/02/2023 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 165.6 F 2.0
4:00 187.6 F 6.4
6:00 207.3 F 12.0
8:00 226.0 F 7.4
10:00 235.5 F 13.5
12:00 241.0 F 11.7
14:00 240.2 F 10.7
16:00 254.4 F 17.8
18:00 266.7 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.7 F 29.5
2:00 271.4 F 28.7
3:00 271.9 F 29.4
4:00 272.5 F 29.3

Min 271.0 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 223.4 F 0.0
10:00 199.8 F 0.0
15:00 197.9 F 0.0
20:00 194.9 F 0.0
25:00 192.6 F 0.0
30:00 190.8 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

11/02/2023 10:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____