
Midmark M11 - v1.0.0
Total Cycles: 1150

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2021 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.6 F 0.0
2:00 208.9 F 3.6
4:00 225.7 F 8.7
6:00 237.4 F 11.1
8:00 232.6 F 10.3
10:00 234.7 F 8.0
12:00 242.3 F 11.7
14:00 245.9 F 13.3
16:00 258.1 F 19.8
18:00 269.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.9
2:00 271.6 F 28.8
3:00 272.1 F 28.6
4:00 271.9 F 29.0

Min 271.1 F 28.3
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 238.6 F 0.0
10:00 223.6 F 0.0
15:00 212.1 F 0.0
20:00 214.0 F 0.0
25:00 202.2 F 0.0
30:00 204.9 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

09/08/2021 01:07 PM
ACCEPTED: X
REJECTED:
Operator: U01