

-----  
Midmark M11 - v1.0.5  
Total Cycles: 781

Name: RITTER2  
Sterilizer ID: V2313228  
02/10/2022 12:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.1 F 0.0  
2:00 184.7 F 3.5  
4:00 204.0 F 9.0  
6:00 220.3 F 8.5  
8:00 232.8 F 10.9  
10:00 233.9 F 9.2  
12:00 237.9 F 9.8  
14:00 249.3 F 15.6  
16:00 261.9 F 22.7  
18:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.4 F 29.5  
2:00 272.1 F 29.8  
3:00 272.1 F 29.4  
4:00 271.9 F 29.6  
  
Min 271.5 F 29.0  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.7 F 0.0  
5:00 218.9 F 0.0  
10:00 236.7 F 0.0  
15:00 222.6 F 0.0  
20:00 218.0 F 0.0  
25:00 218.2 F 0.0  
30:00 216.9 F 0.0  
  
Drying Complete

FILLING: 2:56  
HEATING: 18:17  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:27  
CYCLE COMPLETE

02/10/2022 01:13 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02