

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5057

Name: RITTER 1  
Sterilizer ID: V2178198  
11/05/2024 12:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.2 F 0.0  
2:00 175.1 F 0.4  
4:00 180.7 F 2.9  
6:00 199.8 F 7.0  
8:00 214.6 F 11.2  
10:00 220.8 F 5.3  
12:00 232.1 F 10.3  
14:00 233.1 F 8.3  
16:00 243.1 F 12.8  
18:00 241.6 F 11.2  
20:00 253.8 F 17.2  
22:00 264.3 F 23.5  
24:00 271.7 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.3 F 29.0  
2:00 271.9 F 28.5  
3:00 271.1 F 28.3  
4:00 272.2 F 29.0

Min 271.1 F 28.1  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 205.0 F 0.0  
10:00 203.5 F 0.0  
15:00 200.3 F 0.0  
20:00 201.6 F 0.0  
25:00 208.0 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 24:07  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:02:52  
CYCLE COMPLETE

11/05/2024 01:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_