
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001393
11:20 21/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:58
PRESSURIZING 3:58
STERILIZING 5:12
136.0 C 221kPa 5:12
136.2 C 222kPa 5:42
136.4 C 225kPa 6:12
136.4 C 225kPa 6:42
136.3 C 224kPa 7:12
136.2 C 223kPa 7:42
136.3 C 225kPa 8:12
136.3 C 225kPa 8:42
136.3 C 225kPa 8:42
VENTING 8:42
AIR DRYING 9:37
CYCLE COMPLETE 10:01

Digital Signature #
F1D6079017067C05
