
Midmark M11 - v1.0.0
Total Cycles: 4872

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2024 10:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.6 F 0.0
2:00 178.1 F 0.2
4:00 180.6 F 0.8
6:00 200.5 F 3.6
8:00 214.1 F 7.1
10:00 225.7 F 11.3
12:00 225.7 F 5.9
14:00 237.5 F 10.9
16:00 234.6 F 7.6
18:00 245.0 F 12.2
20:00 245.2 F 12.6
22:00 256.6 F 18.6
24:00 266.6 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.8
2:00 272.3 F 28.5
3:00 271.7 F 28.0
4:00 271.8 F 28.5

Min 271.5 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 220.9 F 0.0
10:00 217.1 F 0.0
15:00 195.6 F 0.0
20:00 196.3 F 0.0
25:00 195.3 F 0.0
30:00 199.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:47
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:04:36
CYCLE COMPLETE

09/06/2024 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____