
Midmark M11 - v1.0.0
Total Cycles: 1638

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2022 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 183.3 F 2.2
4:00 202.9 F 6.0
6:00 215.6 F 10.8
8:00 218.6 F 4.6
10:00 233.3 F 10.3
12:00 232.0 F 7.2
14:00 244.0 F 13.0
16:00 242.9 F 11.4
18:00 255.9 F 18.2
20:00 267.4 F 25.3
22:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.8 F 28.9
2:00 271.6 F 28.8
3:00 272.6 F 28.8
4:00 271.4 F 28.6

Min 271.2 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.8 F 0.0
5:00 226.6 F 0.0
10:00 239.3 F 0.0
15:00 229.1 F 0.0
20:00 224.0 F 0.0
25:00 231.2 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:09
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

02/15/2022 04:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____