

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3962

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2024 10:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.5 F 0.0  
2:00 127.2 F 0.1  
4:00 137.8 F 0.8  
6:00 172.3 F 4.1  
8:00 193.2 F 7.8  
10:00 205.2 F 12.1  
12:00 218.6 F 5.2  
14:00 232.6 F 9.3  
16:00 240.2 F 12.0  
18:00 239.5 F 10.9  
20:00 235.9 F 8.9  
22:00 249.0 F 15.0  
24:00 259.7 F 21.0  
26:00 270.0 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.6  
1:00 272.5 F 29.5  
2:00 271.7 F 28.7  
3:00 271.5 F 29.1  
4:00 272.4 F 29.4

Min 271.0 F 28.5  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
5:00 233.4 F 0.0  
10:00 225.9 F 0.0  
15:00 186.4 F 0.0  
20:00 175.1 F 0.0  
25:00 173.4 F 0.0  
30:00 171.7 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 27:00  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:05:35  
CYCLE COMPLETE

09/20/2024 11:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_