

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2301

Name: RITTER 1  
Sterilizer ID: V2178198  
09/08/2022 06:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.4 F 0.0  
2:00 88.2 F 0.0  
4:00 152.1 F 0.1  
6:00 178.4 F 0.3  
8:00 189.5 F 2.1  
10:00 206.3 F 5.6  
12:00 220.6 F 9.2  
14:00 232.5 F 12.8  
16:00 226.2 F 6.5  
18:00 234.0 F 10.4  
20:00 235.8 F 8.5  
22:00 239.9 F 10.9  
24:00 233.4 F 7.2  
26:00 247.0 F 13.8  
28:00 257.7 F 19.6  
30:00 267.9 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 29.1  
1:00 272.2 F 29.3  
2:00 271.7 F 29.2  
3:00 272.6 F 29.1  
4:00 271.8 F 29.4  
  
Min 271.3 F 29.0  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
5:00 208.4 F 0.0  
10:00 232.5 F 0.0  
15:00 208.2 F 0.0  
20:00 204.4 F 0.0  
25:00 204.8 F 0.0  
30:00 204.3 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 31:45  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:09:50  
CYCLE COMPLETE

09/08/2022 08:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_