
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000168
8:49 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:48
STERILIZING 5:28
136.4 C 226kPa 5:28
136.2 C 223kPa 5:58
136.4 C 226kPa 6:28
136.4 C 226kPa 6:58
136.4 C 226kPa 7:28
136.4 C 225kPa 7:58
136.4 C 225kPa 8:28
136.4 C 226kPa 8:58
136.4 C 226kPa 8:58
VENTING 8:58
AIR DRYING 10:02
CYCLE COMPLETE 25:02

Digital Signature #
047DFAB6A01BCFAC
